STANDARD PLATE AND COLIFORM COUNT AGAR POUR PLATE METHODS IMS #2 (SPC), IMS #21 (CPC)

[Unless otherwise stated all tolerances are ±5%]

SAMPLES

1.			cory Sample Requirements (see CP items 33 & 34) ibitor testing requirements, refer to Section 6 of the PMO]	
			MEDIA PREPARATION	
2.	Med	dia P	reparation (reference agars/broth from CP items 14, 27, 28 & 29)	
	a.	Ter	mperature Control (TC) used for each test agar type	
		1.	Contains agar identical to type and volume being used	
		2.	In container identical in size and volume to that being used	
		3.	Undergoes same heat treatment and cooling as test agar	
	b.	Pla	te Count Agar or Standard Methods Agar (PCA or SMA)	
		1.	Prepare and sterilize agar for sample series and all controls	
		2.	OR use previously prepared/stored agar; melt agar quickly in boiling water or flowing steam; not under pressure	
		3.	Do not melt agar more than once	
		4.	Promptly place in a circulating water bath to temper, hold melted agar at 45 ±1°C	
		5.	Record agar temperature with other control information	
		6.	Agar should be discarded if not used within 3 hours after tempering	
		7.	Avoid prolonged exposure to high temperatures during and after melting; establish lab protocol	
	c.	Vio	let Red Bile Agar (VRB)	
		1.	Boil for at least 1 min, but no more than 2 min. Do not autoclave.	
		2.	Promptly place in a circulating water bath to temper; hold melted agar at 45±1°C	
		3.	Record agar temperature with other control information	

		4.	Agar should be discarded if not used within 3 hours after tempering	
	d.	Brill	liant Green Lactose Bile Broth (BGB)	
		1.	Examine Durham/fermentation tubes for presence of air bubbles	
		2.	If air bubbles cannot be removed from tubes; DO NOT USE	
			PROCEDURE	
3.	Woi	rk Ar	ea	
	a.	Leve	rel plating bench not in direct sunlight	
	b.	San	nitize immediately before start of plating	
4.	Sele	ectin	g Dilutions	
	a.	Star	ndard Plate Count (SPC)	
		1.	Plate two decimal dilutions per sample	
		2.	Select dilutions that would be expected to yield one plate with 25-250 colonies	
			a. Raw milk is normally diluted to 1:100 and 1:1000	
			b. Finished products are normally diluted to 1:10 and 1:100	
		3.	SPC not performed on cultured or acidified products	
	b.	Coli	iform Plate Count (CPC)	
		1.	For pasteurized fluid milk samples, 1 mL direct and/or decimal dilutions as appropriate	
		2.	For samples other than milk (item 11) and dry milk products (item 12) distribute 10 mL of a 1:10 dilution among three plates	
5.	lder	ntifyii	ng Plates	
	a.	with	ect number of samples in any series so that all will be plated nin 20 min (pref. ≤ 10) after diluting first sample and pouring the plate in the series	
	b.	Lab	pel each plate with sample or control identification and dilution	
	C.	Arra	ange plates in order before preparation of dilutions	

CONTROLS

6.	Controls (AM and PM)				
	a.	Check sterility of dilution blanks, agar, Petri dishes, and pipets/tips used for each group of samples			
	b.	Expose a poured plate to air with cover completely removed during plating for 15 min; timer used			
		The air control plate must be the first plate poured immediately before samples are shaken and must be located such that it is in the area of the plating activity (not off to the side)			
		2. After incubation, air plate(s) shall contain ≤15 colonies			
		Take and record corrective actions for air control plate(s) with >15 colonies			
	c.	Maintain records			
	d.	Include information on bench sheet, work sheet or report sheet(s)			
		DILUTING SAMPLES			
7.	San	Sample Agitation			
	When appropriate, wipe top of unopened containers with sterile, ethyl alcohol-saturated cloth				
	b.	Before removal of any portion or sub-samples, thoroughly mix contents of each container			
		 Mix raw milk sample(s) by shaking 25 times in 7 sec with a 1 ft movement (containers approx ¾ full) 			
		2. Mix retail milk samples by inverting containers top to bottom, then bottom to top (a complete half circle or 180 degrees) without pausing, 25 times			
	c.	Remove test portion within 3 min of sample agitation			
8.	Dilu	ution Agitation			
	 a. Before removal of any portion, shake each dilution bottle 25 times in 7 sec with a 1 ft movement 				
	b.	Remove test portion within 3 min of dilution agitation			

	C.	Mechanical shakers may be used only if a laboratory provides validation data on a specific unit. Data must pass validation criteria
		PLATING
9.	San	nple & Dilution Measurements, Pipets
	a.	Use separate sterile pipets for the initial transfers from each container, adjusting pipets in pipet container without touching the pipets
	b.	Do not drag pipet tip over exposed exterior of pipets in pipet container
	C.	Do not drag pipet across lip or neck of sample container or dilution blank
	d.	Insert pipet not more than 2.5 cm (1") below sample surface or dilution surface (avoid foam and bubbles)
	e.	Using pipet aid, draw test portion above pipet graduation mark and remove pipet from liquid (mouth pipetting not permitted)
	f.	Adjust test volume to mark with lower side of pipet:
		In contact with inside of sample container (above the sample surface)
		Or, in contact with inside of dilution blank neck or area above buffer on straight-walled container
		Ensure excess liquid does not adhere when pipet is removed from the sample container or dilution blank
	g.	For dilutions, dispense test portion to dilution blank (with lower side of pipet in contact with neck of dilution blank, or area above buffer on straight-walled containers) with column drain of 2-4 sec
	h.	Gently lift cover of Petri dish just high enough to insert pipet
	i.	Hold pipet at approximately a 45° angle with tip touching dish
	j.	Release sample or dilution portion to Petri dish (with lower side of pipet in contact with plate) with column drain of 2-4 sec
		Using pipet aid, blow out the last drop of undiluted sample, away from main part of sample
		On diluted samples, touch pipet tip once against dry spot on dish bottom
		a. When depositing 0.1mL, do not re-touch to dry area

	k.	Discard pipets into disinfectant OR dispose into biohazard bags or containers to be sterilized (using this method of disposal does not require placing into disinfectant first)		
10.		Sample & Dilution Measurements, Pipettors [for electronic pipettors, follow manufacturer instructions] Mechanical Electronic		
	a.	Each day before use, vigorously depress plunger 10x to redistribute lubrication and assure smooth operation (mechanical pipettors)		
	b.	Before each use, examine pipettor to assure that no liquid is expelled from the pipettor nose-cone (contaminated), if fouling is detected do not use until cleaned as per manufacturer recommendation		
	c.	Use separate sterile tip for the initial transfers from each container		
	d.	Depress plunger to first stop (mechanical pipettors)		
	e.	Do not drag tip/barrel across lip or neck of sample container or dilution blank, and do not allow pipettor barrel within sample container		
	f.	Insert tip approximately 0.5-1.0 cm below sample or dilution surface (avoid foam and bubbles)		
	g.	With pipettor vertical, slowly and completely release plunger on mechanical pipettor; do not lay pipettor down once sample is drawn up, use vertical rack or charging stand if necessary		
	h.	Touch off lower side of tip:		
		To inside of sample container above the sample surface, excess liquid not adhering to tip		
		Or to the inside of dilution blank neck or area above buffer on straight-walled containers, excess liquid not adhering to tip		
		a. For dilutions, hold pipettor nearly vertical with lower side of tip touching neck of dilution blank (or area above buffer on straight-walled containers), dispense test portion to blank by slowly depressing plunger to stop (mechanical pipettor)		
		3. For two (2) stop pipettors, depress plunger to second stop with tip remaining in contact with dilution blank		
	i.	Gently lift cover of Petri dish just high enough to insert tip with pipettor approximately vertical to dish		
		Release sample or dilution portion onto plate with tip slightly above but not in contact with the plate by slowly depressing plunger completely		

		a.	For two (2) stop pipettors, depress plunger to second stop	
		b.	Do not touch off pipettor tip(s) to Petri dish	
		C.	Optionally, deposit samples with pipettor capable of making a 1:10 dilution in the tip	
	j.	containe	tips into disinfectant OR dispose into biohazard bags or ers to be sterilized, (using this method of disposal does not placing into disinfectant first)	
11.	San	nples Oth	ner than Milk	
	a.	Weigh 1	1 g aseptically into a 99 mL dilution blank heated to 40-45°C	
12.	Dry	Milk Pro	duct Samples	
	a.	Weigh 1	1 g aseptically into a 99 mL dilution blank heated to 40-45°C	
		1. Use	e standard dilution blank	
			2.0 % sodium citrate blank (pH<8.0) for relatively insoluble mple (e.g. whey)	
	b.	Wet sam	nple completely with gentle inversions	
	C.		c a minimum of 2 min; shake 25 times in 7 sec with a ement; use within 3 min of agitation	
13.	Pou	ring Aga	ır	
	a.	each pla agar wei actions (spensing test portions, promptly pour 10-12 mL of agar into ate of series, or 12-15 mL for > 1 mL portion/plate or where ight loss is a problem that cannot be corrected by other (Documentation must be kept to indicate that this is a practice; amount poured to match agar weight loss test)	
		1. Lift	cover of Petri dish just high enough to pour agar	
			er agar is poured, thoroughly and evenly mix agar and test rtion in Petri dish(es)	
		a.	Agar solidification to occur within 10 min	
		b.	Do not stack plates prior to solidification	
	b.	For dry r	milk product sample(s), overlay plate with 3-5 mL PCA or SMA	
	c.	For colif	orm count, overlay plate with 3-4 mL VRB	

INCUBATION

14.	Incubating Plates (see CP item 15)			
	a.	Stack plates (upside down) no more than 6 high and incubate within 10 min of agar solidification		
	b.	Place stacks to ensure adequate air flow		
	C.	Incubate SPC plates at 32±1°C for 48±3 hours (dry milk products for 72±3 hours)		
	d.	Incubate Coliform plates at 32±1°C for 24±2 hours		
		COUNTING COLONIES		
15.	Cou	nting Aids (see CP items 16 and 17)		
	a.	Count colonies with Quebec dark-field model or equivalent with satisfactory grid plate (CP item 16)		
	b.	Hand tally (see CP item 17)		
16.	Cou	nting, Recording and Computing SPC		
	a.	After incubation, count all colonies on selected plates		
	b.	Where impossible to count at once, store plates at 0.0-4.5°C for not longer than 24 hr (avoid as a routine practice)		
	c.	Record results of sterility and control tests		
	d.	Record dilutions used and number of colonies on each plate counted		
	e.	When possible, select spreader free plates with 25-250 colonies and count all colonies including those of pinpoint size		
		Use higher magnification if necessary to distinguish colonies from foreign matter		
		Examine edge of Petri dishes for colonies		
	f.	If consecutive plates yield 25-250 colonies, count all colonies on plates from both dilutions		
	g.	Spreaders		
		Count colonies on representative portion only when colonies are well distributed and area covered or repressed does not exceed 25% of plate		

	2.	Do not count if repressed growth area > 25% of plate area	
	3.	When spreaders must be counted, count each as a single colony	
	4.	Count chains/spreaders from separate sources as separate colonies	
	5.	If 5% of plates are more than 25% covered by spreaders, take Immediate steps to eliminate and resolve problem	
h.		ere is no plate yielding 25-250 colonies, use plate having nearest 50 colonies	
i.	If pla	ates from all dilutions exceed 250 colonies, estimate counts as ws	
	1.	Count colonies in portions representative of distribution and estimate total	
	2.	Where there are < 10 colonies/sq. cm, count colonies in 12 squares, selecting 6 consecutive squares horizontally across the plate and six consecutive squares at right angles	
	3.	When there are 10 or more colonies/sq. cm, count 4 random representative squares	
	4.	Multiply average number colonies/sq. cm by area of plate in sq cm	
j.		ates from all dilutions yield < 25 colonies each, record actual	
k.	If all	plates from a sample show no colonies, record count as 0	
l.		iply number of colonies (or estimated number if necessary) ne reciprocal of the dilution	
	1.	If consecutive dilutions yield 25-250 colonies, compute count using formula below	
		$N = \sum C/[(1 \times n1) + (0.1 \times n2)]d$	

number of colonies per milliliter or gram

sum of all colonies on all plates counted

number of plates in lower dilution counted

number of plates in next highest dilution counted

dilution from which the first counts were obtained

Where,

N =

 $\Sigma C =$

n1 =

n2 =

d

24,727 [25,000 (reported)] In the NCIMS Program the denominator will always be 0.11 Note: for 1:10 dilutions and 0.011 for 1:100 dilutions 17. Counting, Recording and Computing CPC After incubation, count all colonies on selected plates a. Where impossible to count at once, store plates at 0.0-4.5°C for not b. longer than 24 hr (avoid as a routine practice) Confirmation of colonies C. 1. Pick 10% up to 10 representative colonies per plate with relative percentages of each colony type and inoculate into BGB; incubate for 48±3 hours at 32±1°C 2. Gas production at any time during the incubation is considered a confirmed test 3. Record the number of picked colonies and the number of colonies that produced gas (if necessary calculate % confirmed and multiply by total number of colonies) If no colonies appear on plate(s), record count as 0 d. If there are 1-154 colonies on a plate, record number counted e. f. If >154 colonies develop on the highest dilution plate, record number as >150 When multiple plates of a dilution are used, sum counts of plates g. Multiply number of colonies (or estimated number if necessary) h. by the reciprocal of the dilution **Identifying Counting Errors** Perform monthly counting for SPC a. 1. With 3 or more analysts, use the RpSm method (see current SMEDP); maintain records

1:1.000 = 28 colonies

Example: 1:100 = 244 colonies

272/[1.1]0.01 272/0.011

 $(244 + 28)/[(1 \times 1) + (0.1 \times 1)]0.01$

Ν

		2.	With two analysts, comparative counts agree within ≤ 10% of one another; maintain records	
		3.	With only one analyst, replicate counts agree within ≤8% of one another; maintain records	
			REPORTING	
19.	[Wh	en sa nts a	ng (see CP item 34.b.2.d) amples are demonstrated to contain inhibitors, no bacteria are reported; report as positive for inhibitors or growth s (GI)]	
	a.	SPC		
		1.	Report computed count as Standard Plate Count/mL or /g (SPC/mL or SPC/g) when taken from plate(s) in the 25-250 range	
		2.	Report SPC plate counts of 0 to 24 as < 25 times the reciprocal of the dilution and report as Estimated SPC (ESPC)	
		3.	When colonies on SPC plates exceed 100/sq. cm, compute count by multiplying 100 x dilution factor x area of plate in sq. cm and report as > computed count Estimated (ESPC)	
		4.	If computed counts from SPC plates are >250, report as Estimated SPC (ESPC)	
	b.	CPC		
		1.	Report count as Coliform Plate Count (confirmed)/mL or /g when taken from plate(s) in the 1-154 range (CPC/mL)	
		2.	If no colonies appear on coliform plates, report as < 1 times the reciprocal of the dilution and report as Estimated (ECPC)	
		3.	Counts from coliform plates > 154 are reported as > 150 Estimated Coliform Count (ECPC)	
		4.	If for any reason, an entire plate is not counted, the computed count is reported as Estimated (ECPC)	
	c.	Rep	oort only first two left-hand digits	
		1.	If the third digit is 5 round the second number using the following rules	
			a. When the second digit is odd round up (odd up, 235 to 240)	
			b. When the second digit is even round down (even down, 225 to 220)	

d.	If all plates from a sample have excessive spreader growth, report as spreaders (SPR)	
e.	If a laboratory accident renders a plate uncountable, report as laboratory accident (LA)	