

Facility/Laboratory Name: _____

INCUBATOR TEMPERATURE LOG

OFFICIAL RECORDS FOR THE PERIOD _____ TO _____

Make/ Model: _____ Serial Number: _____ Unit Id#/ Identification#: _____ Temperature Range: _____

Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
AM																															
PM																															
Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
AM																															
PM																															
Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
AM																															
PM																															

1. Sufficient size to prevent crowding of interior.
2. Shelves placed to assure uniformity.
3. Chamber temperatures measured by not less than two thermometers with bulbs immersed in liquid (in sealed containers).
4. Temperature (corrected) recorded from each thermometer twice daily (AM and PM).
5. Record temperature to the nearest 0.1°C along with the analyst's initials. Example

45.5°C
MH